

Sterilization certified in line with

EN ISO 13485:2016 / EN ISO 11135:2014



MEDICOPLAST
International GmbH

Confirmation of Sterilization

Method: VA_S_1

Sterilization date:	2020-03-09
Sterilization lot:	18099
Customer:	MT.Derm GmbH
Your delivery note:	20-EB0981, 20-EB0982
Your requisition no.:	20-EB0983, 20-EB1034, 20-EB1037, 20-EB1038, 20-EB1041, 20-EB1053, 20-EB1056, 20-EB1057, 20-EB1058, 20-EB1067
Our delivery note:	249604 from 2020-03-12
Item:	9900031 Sterilisation Pal. MP01 1d NK Pr.70
Quantity:	2 Pal.

	Nominal values	Actual values
<u>Incoming time:</u>	min. 12 h	59 h
<u>Preconditioning</u>		
Dwell:	min. 16 h	46,25 h
Conditions:	35-45°C 50-80% rel.humidity	35-45°C 50-80% rel.humidity
<u>Sterilization</u>		
Cycle:	MP1	
Temperature:	40 ±3 °C	40 °C
Prevacuum:	120 ±10 mbar abs	127 mbar abs
Humidity:	50-80% rel.humidity	63,8% rel.humidity
Dwell time:	3 h	3 h
EO mass:	16-32 kg	25,5 kg
Air washes:	min. 3	3
<u>Aeration</u>		
Dwell:	min. 1 Days	2 Days
Condition:	35-45 °C	35-45 °C


Remarks: The evaluation from the BI's is in process.

The physical parameters were in accordance to the specifications.

No growth was observed for the microbiological inprocess control with biological indicators (Bacillus atrophaeus, 3M Attest).

The sterilization was conducted in line with the customer validation parameter.

Approval

Signature QS:  Date: 12/3/20

*Freigegeben
Gosciniak
13.03.2020
OK GEPRÜFT*
