

Peak Needles

7410 Coca Cola Dr. unit 107
Hanover MD 20176
USA

Guangdong Heshan Sentai Plastic Factory Co., Ltd.

Add:No.WenMin,Lane,DF93,HenSan,District,Jiangmen,City,Guangdong,Province.P.R.China

STERILIZATION CERTIFICATE

TYPE OF PRODUCT SUBJECT OF STERILIZATION:TRITON CARTRIDGES
NEEDLE

TYPE OF STERILIZATION :EO GAS

ITEM NO:TATTOO NEEDLES

DATE OF STERILIZATION: Mar 2022

LOT NO: 032022

We hereby certify that the above mentioned item which were
sterilized by EO GAS on Mar,2022 With the lot NO.032022

*The attachment is the detailed parameter report of EO GAS
sterilization



Peak Needles

7410 Coca Cola Dr. unit 107
Hanover MD 20176
USA

Guangdong Heshan Sentai Plastic Factory Co., Ltd.

Add: No. WenMin, Lane, DF93, HenSan, District, Jiangmen, City, Guangdong, Province. P. R. China

E0-Sterilization Record Per Load

Date: 2022-03-25				Order: 2		Sterilization No: 032022		
Article Name	Production No	Specification	Quantity	Required Operation During Sterilization Cycle				
					Required	Actual		
TRITON Cartridge needle	032022-1	51X36X21 cm	1-80	Tr. C	49-54	50		
TRITON Cartridge needle	032022-2	51X36X21 cm	81-143	Pressure (KPa)	68- 71 P	72 P		
				Humidity (%RH)	30-85%	70		
				Exposure time (min)	12:00	12:00		
				Speed of Air Removal (Kpa/min)	1	2.5		
				Speed of Aeration (Kpa/min)	//	//		
				Weight of EO used Before (kg)	//	90		
				Weight of EO used Afer (kg ⁹)	//	60		
				Weight of EO used (kg)	8	8		
				Pressure Changed (KPa)	s: // E: //	S-69.8 E:34.5		
						2 FO 1		



032022-1					
Required Operation	Start Time	End Time	Actual Time	Sterilization Information	Release
Conditioning	17:00	19:00	120	Nomal	
Air Removal	19:00	19:20	20	Nomal	
Pressure Time	19:20	20:00	40	Nomal	
Eo ijection	20:00	20:30	30	Nomal	
Exposure time	20:30	8:30	720	Nomal	
Eo removal	8:30	9:30	60		
Flushing	9:30	12:00	150		
032022-2					
Required Operation	Start Time	End Time	Actual Time	Sterilization Information	Release
Conditioning	13:00	15:00	120	Nomal	
Air Removal	15:00	15:20	20	Nomal	
Pressure Time	15:20	16:00	40	Nomal	
Eo ijection	16:00	16:30	30	Nomal	
Exposure time	16:30	4:30	720	Nomal	
Eo removal	7:00	7:30	60		
Flushing	7:30	10:00	150		
					2 FO 2

