



Bemis Two-Gallon Sharps Container

Model 102

Instructions for Use

1. Align front of lid (lid opening) to front of base (biohazard label). Press lid firmly to base ensuring that the three snaps in front and back and two on each side are engaged.
The parts will not snap together if not correctly aligned. Be sure the handles on both lid and base meet.
2. With one hand, drop sharps into container. Check fill level regularly. Replace with new container when fill level is reached.
3. Needle Key Feature
With container mounted in bracket, insert needle completely into slot and slide needle holder to right or left, engaging the the needle hub into key. Rotate counter-clockwise until needle is unscrewed from holder. With blunt tip of syringe/holder, nudge the hub back to the center hole to release needle into container.
The needle key feature must only be used when container is secured in wall bracket.
4. Temporary Closure
Bring lid forward and press down on two front snaps. Do not insert tab into opening. Use tab to reopen lid.
5. Final Closure
Bring lid forward, insert tab into slot, and press down on two front snaps to lock lid closed.
6. Disposal
Containers may be autoclaved or incinerated. Dispose of according to hospital policy.

For Use with Wall Bracket

Model 435 020

Slide container into bracket so that sides of bracket fit indentations of unit. Push down until it clicks. Pull up on container to be sure it is locked in place.

To remove container from bracket, insert plastic key (Part # 440 020) into slots at bottom of bracket and push up until container is released.

